



Continuation of Substance of Interview including description of the general nature of what was discussed: Indicated to Ms. Gowen that the working examples in the specification do not provide support or data excluding use of low ionic strength diluent solution (LISS) in admixture with the antigen positive RBCs to reduce time for agglutination formation in immunohematology testing in combination with the parameters of temperature, time, and continuous agitation, as claimed. Suggested to Ms. Gowen to amend claim 1 to incorporate the LISS in admixture with the antigen positive RBCs. Ms. Gowen gave the Office authorization to amend claim 1 to recite, "antigen positive red blood cells admixed with low ionic strength diluent" and to cancel claims 8 and 10-24 in an Examiner's Amendment.